

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 14 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130121 ✓
Name	John Pennell	ID#	6260 Date 8/13/09 ✓
Agency	Fairbanks Airport Police	Phone #	474-2555
Instrument Location	6450 Airport Way Ste. 7 Fairbanks, AK 99709 ✓		
Alco S/N	57447 ✓	Target Value	.079 ✓
		High Pressure	425 psi ✓
Alco Test Values	.080 ✓	.080 ✓	
	1 st Alco	2 nd Alco	
Signature	John Pennell ✓		
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of August, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

AUGUST 13, 2009 ✓ *COS 8/13/09*

OPERATOR'S NAME:
PERNELL/JOHN/T
OPERATOR'S NUMBER: 6260
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: UOC
DEPT/AGENCY: FBK6
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 57447 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.94 in ✓
ALCO TARGET .079 06:53
BLANK TEST .000 06:54
INTERNAL STANDARD VERIFIED 06:54
ALCO TV 29.94 in .000 06:55
BLANK TEST .000 ✓ 06:56
SUBJECT SAMPLE .000 06:56
BLANK TEST .000 ✓ 06:57
ALCO TV 29.94 in .000 06:57
BLANK TEST .000 06:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

AUGUST 13, 2009 ✓
TIME 06:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 44c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~